

## **What to do when your municipal water supplier issues a "boil water advisory"**

- **What is a "Boil Water Advisory?"**

A Boil Water Advisory (BWA) is issued to protect the health of the community from the possibility of waterborne infectious agents. A BWA is issued only after careful consideration among representatives from public health, regulatory agencies and municipal departments after positive tests (e.g., positive samples for fecal coliforms, changes in turbidity measurements), or line breaks.

- **What should I do when under a BWA?**

For personal use, create a supply of water for cooking, drinking and tooth brushing by bringing water to a rolling boil for 1 minute. Timing starts when the water starts to bubble. Cool the water then place in clean containers for use or refrigerate.

Hot (not boiled) soapy water can be used for dishwashing and kitchen / bathroom surface cleaning. As a precaution, add one tablespoon of bleach per gallon. Laundry water does not need to be treated. Unless specifically listed on your notice, water for showering does not need to be treated.

- **Can we dialyze patients during a BWA?**

Yes, if the water treatment components in use are sufficient to remove or destroy bacteria Reverse Osmosis (RO) will protect the product water from having microbial contamination. Deionization (DI) does not remove or destroy bacteria, so if DI is being used as the main water treatment (rather than RO), you will need a submicron or endotoxin/ultrafilter downstream of the DI. If an Ultraviolet (UV) irradiator is used, the filter should be located after the UV irradiator. Close monitoring of the resistivity of the product water will be needed to detect any decrease in quality. Also consider weekly microbial assessment of the product water during the BWA.

Keep in close contact with the municipal water supplier because they may choose to "shock" treat (hyperchlorinate) their distribution system to bring it back into compliance with the acceptable standards for drinking water. If the city "shocks" their water system, you may see chlorine/chloramine break through. Review your testing procedures with staff and alert them to be vigilant for potential break through so that patients will be protected from exposure to chlorine/chloramine.

Prepared by Matthew Arduino MS, DrPH, with CDC, and Glenda M. Payne, RN, MS, CNN, with CMS. The opinions and assertions presented here are the private views of the authors and are not to be construed as official or reflecting the views of the Centers for Disease Control or the Centers for Medicare & Medicaid Services.