



The Florida ESRD Network

HACH Expanded Recall Notification SteriChek “Sensitive” Total Chloramines and Residual Chlorine Reagent Strips

Hach Company is expanding its voluntary recall of SteriChek “Sensitive” Total Chloramines and Residual Chlorine Reagent Strips that may have been shipped on or after August 24, 2006.

The recall is being conducted because further investigation has shown that there have been other lots that have not satisfied Hach’s QC standards. In particular, testing has shown inconsistencies in identifying levels of total chlorine and free chlorine. Such inconsistencies may result in inaccurately high or low readings. All products and lot numbers identified on the attached table that may have been shipped on or after August 24, 2006 are being recalled. Please review this table to determine if you have any of these lot numbers in your inventory.

Information for immediate return of recalled products to Hach Company:

1. First contact Customer Service at 1-800-548-4381 ext. 3560 to receive a return authorization.
2. Acquire return labels from Hach Company
3. Affix the labels to the product
4. Send the product to the following address:

**Hach Company
100 Dayton Rd
Ames, IA 50010**

Complete the attached Return Form and send to the above address. You may also fax the form to 515-956-3820.

This recall is being conducted to the end user level; therefore, distributors may be contacting customers to return the products listed in the attached table.

The Florida ESRD Network (Network 7) is providing this fax blast as a technical assistance activity for the Florida renal community.

FMQAI: The Florida ESRD Network
5201 West Kennedy Boulevard, Suite 900 • Tampa, Florida 33609
813-383-1530 • 813-354-1514 Fax



MINNTECH
14605 28TH AVE NORTH
PLYMOUTH, MN 55447

RETURN FORM

SteriChek "Sensitive" Total Chloramines and Residual Chlorine
Expanded RECALL

<u>Product Description, Product Code</u>	<u>Lot</u>	<u>Quantity</u>
SteriChek Total Chloramines and Residual Chlorine, 811909 & 812144	_____	_____
Sensitive Total Chlorine Reagent Strips, 821972	_____	_____
Please use additional lines to record product description & code, lot and qty.		
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____

* Please indicate the quantity/quantities of SteriChek "Sensitive" Total Chloramines and Residual Chlorine returned to Hach Company.

We have checked our inventory and were unable to locate the above-mentioned product.

Contact Person (Please Print)

Signature

Date

Please return by fax or mail to:

Hach Company
100 Dayton Road
Ames IA 50010
Attention: Diane Draper
Fax Number: 515-956-3820